

Declaration of conformity

We declare under our responsibility, that the following product:

Model: ScanSpeed 406 Low Speed Centrifuge

to which this declaration relates is in conformity with the following standard(s), directives or other normative document(s):

In compliance with:

EN 61010-1 - Safety requirements for electrical equipment for measurement, control and laboratory use - General requirements

EN 61010-2-020 - Safety requirements for electrical equipment, control and laboratory use - Particular requirements for laboratory centrifuges

EN 61000-6-1 - Electromagnetic compatibility - Generic immunity/emission standard

EN ISO 11201 – Acoustics – Noise emitted by machinery and equipment

Following the provisions of:

2006/42/EC - Machinery Directive, as amended

2006/95/EC - Low Voltage Directive, as amended

2004/108/EC - EMC Directive, as amended

2011/65/EU - RoHS Directive

Lynge, September 2014

Rasmus Sørensen QA Manager

Resours Soras

CE

LaboGene ApS, Industrivej 6-8, Vassingerød, 3540 Lynge, Denmark